

# Rivoflavin(Vit B2) Analysis

## 1. Analytical Condition

### YL 9100 HPLC

Mobile Phase : 10mM NaH<sub>2</sub>PO<sub>4</sub>(pH=5.5):MeOH=65:35, (mix in a bottle)

Flow Rate : 1.0ml/min

Column Oven : 40 °C

Detector : UV/Vis 254nm

Column : C18 (4.6\*250 mm 5 um)

Injection Volumn : 50ul

## 2. Chromatogram

