

Tocopherol Acetate Analysis

1. Analytical Condition

YL 9100 HPLC

Mobile Phase : Methanol : Water = 49:1 (mix in a bottle)

Flow Rate : 1.0ml/min

Column Oven : 25°C

Detector : UV/Vis 284nm

Column : Xbridge C18 (4.6*150 mm 5 um)

Injection Volumn : 20ul sample loop

2. Chromatogram

