

Vitamin D₂ Analysis In Tablet

1. Analytical Condition

YL 9100 HPLC

Column : C18 (1.5X250mm, 5 μ m)

Detector : UV/Vis Wavelength : 280nm

Mobile phase : ACN:MeOH=75:25

Flow rate : 0.2 ml/min

Injection volume : 30 μ l

2. Chromatogram

