

Vitamin E Analysis

1. Analytical Condition

YL 9100 HPLC

Mobile Phase : Methanol

Flow Rate : 1.0ml/min

Column Oven : 35°C

Detector : UV/Vis 280nm

Column : C18 (4.6*250mm, 5 μ m)

Injection Volumn : 20ul sample loop

2. Chromatogram

