

General Sugars Analysis

1. Analytical Condition

YL 9100 Series HPLC

Column : Carbohydrate(4.6mm*25cm)

Detector : RID- 35°C

Injection Volume : 20 μ l

Mobile Phase : ACN :Water =75:25 (1 bottle mix)

Flow Rate : 1-1.5ml/min

Oven Temperature : 35°C

2. Chromatogram

