

Glucose/Arabinose Analysis

1. Analytical Condition

YL 9100 Series

Mobile phase : Water

Flow rate : 0.4 mL

Column oven : 80°C

Detector : RI 40 °C

Column : KS-802(7.8*300 mm)

2. Chromatogram

