

# Glucosamine Analysis

## 1. Analytical Condition

### YL 9100 Series HPLC

Mobile Phase : ACN:Water=70:30 (Isocratic mode)

Flow Rate : 1.5ml/min

Column Oven : 35°C

Detector : ELSD – Range ¼, gas 2L/min, Temp 60°C

Column : Carbohydrate (3.9\*300mm)

Injection Volumn : 20ul sample loop

## 2. Chromatogram

