

# Ribose, Xylose, Xylitol, Mannitol Analysis

## 1. Analytical Condition

### YL 9100 Series HPLC

Mobile Phase : ACN:Water=85:15(Isocratic mode)  
Flow Rate : 1.5ml/min  
Column Oven : 35°C  
Detector : ELSD – Range ¼, gas 2L/min, Temp 60°C  
Column : Carbohydrate (4.6\*250mm)  
Injection Volumn : 20ul sample loop

## 2. Chromatogram

