



# Sorbitol, Sorbitan, iso-Sorbitide Analysis

## 1. Analytical Condition

### YL 9100 Series HPLC

Mobile Phase : ACN:Water=80:20 (mix it in a bottle)

Flow Rate : 1ml/min

Column Oven : 35°C

Detector : RI Det, 35°C

Column : Carbohydrate (4.6\*250mm, 5um)

Injection Volumn : 20ul sample loop

## 2. Chromatogram

