

# General Sugars Analysis

## 1. Analytical Condition

### YL 9100 Series HPLC

Column : Sugar SP0810 (8.0mm\*30cm)

Detector : RID- 35°C

Injection Volume : 20 $\mu$ l

Mobile Phase : Water 100%

Flow Rate : 0.8ml/min

Oven Temperature : 80°C

## 2. Chromatogram

