

Xylose, Fructose, Mannitol, Glucose Analysis

1. Analytical Condition

YL 9100 Series HPLC

Mobile Phase : ACN:Water=85:15(Isocratic mode)

Flow Rate : 1.5ml/min

Column Oven : 35°C

Detector : ELSD – Range ¼, gas 2L/min, Temp 60°C

Column : Carbohydrate (4.6*250mm)

Injection Volumn : 20ul sample loop

2. Chromatogram

